

CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Millstone Medical Testing - CT, LLC 35 Griffin Rd South Bloomfield, CT. 06002

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

TESTING

This certificate is valid only when accompanied by a current scope of accreditation document. The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 18 October 2026 Certificate Number: AT-3317





SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Millstone Medical Testing - CT, LLC

35 Griffin Rd South Bloomfield, CT. 06002

Victoria Hughes 508-728-5608 victoria.hughes@millstonemedical.com

TESTING

Valid to: October 18, 2026 Certificate Number: AT-3317

Biological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Sterility Test for Biological Indicators/Paper Strips	LSPEC M100 USP<55> ANSI/AAMI/ISO 11138	Medical Devices	Media Incubators
Product Sterility	LSPEC M150 USP <71>, ISO 11135, ISO 11138, ANSI/AAMI/ISO 11737-2, ANSI/AAMI/ISO 11137-2, ANSI/AAMI/ISO 13004	Medical Devices	ISO 7 Clean Room ISO 5 Laminar Flow Hood Media Incubators
Total Bioburden	LSPEC M200 LSPEC M200P ANSI/AAMI/ISO 11737-1	Medical Devices	Media ISO 5 Laminar Flow Hood Incubators
Bacterial Endotoxin Gel clot method	LSPEC M407 ANSI/AAMI ST72 USP <161> USP <85>	Medical Devices	Lysate CSE Heat Block

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-3317.

Jason Stine, Vice President

Version 001 Issued: October 18, 2024

www.anab.org

