



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Millstone Medical Testing-RI, LLC**  
**1300 Main Street**  
**West Warwick, RI 02893**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 27 August 2027

Certificate Number: AT-3412



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

### Millstone Medical Testing – RI, LLC

1300 Main Street,  
West Warwick, RI 02893

Victoria Hughes [victoria.hughes@millstonemedical.com](mailto:victoria.hughes@millstonemedical.com)

Susan Messier [susan.messier@millstonemedical.com](mailto:susan.messier@millstonemedical.com)

### TESTING

ISO/IEC 17025 Accreditation Granted: **27 August 2025**

Certificate Number: **AT-3412** Certificate Expiry Date: **27 August 2027**

#### Microbiological


Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Bacterial Endotoxin (Gel-clot)	BE-001 USP <85>, USP <161>, ANSI/AAMI ST72	Medical Devices, Pharmaceuticals	Depyrogenation Oven Heat block
Bacterial Endotoxin (Kinetic Turbidimetric, Kinetic Chromogenic)	BE-005 USP <85>, USP <161>, ANSI/AAMI ST72	Medical Devices, Pharmaceuticals, Tissue	Depyrogenation Oven Microplate Reader
Bioburden	B-001, B-008 ANSI/AAMI/ISO 11737-1	Textiles, Cosmetics, Medical Devices, Tissues, Pharmaceuticals	ISO Class 5 LFH Incubators
Biological Indicator Sterility, Product Sterility	T-001, USP <71>, ISO 11135, ISO 11138, ANSI/AAMI/ISO 11737-2, ANSI/AAMI/ISO 11137-2, ANSI/AAMI/ISO 13004, ANSI/AAMI/ISO TIR40, ANSI/AAMI/ISO TIR37	Biological Indicators, Medical Devices, Pharmaceuticals, Tissue, Biologics, Laboratory Products	ISO Class 5 LFH Incubators

**Chemical**

Test Method	Test Specification(s)	Range	Comments
EO Residual Testing	RES-01 ANSI/AAMI/ISO 10993-7	Medical Devices	Incubator Gas Chromatograph

**Sample Collection**

Test Method	Test Specification(s)	Range	Comments
Environmental Monitoring- Onsite monitoring of Client Cleanrooms	B-007 ISO 14644-1, -2, -3, ISO 14698, USP <1116>	Cleanrooms - Nonviable Air, Viable Air, Viable Surface	Incubator Nonviable Air Sampler Viable Air Sampler



Jason Stine, Vice President