



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Millstone Medical Testing - CT, LLC**  
**35 Griffin Rd South**  
**Bloomfield, CT. 06002**

Fulfills the requirements of

**ISO/IEC 17025:2017**

In the field of

**TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 18 October 2026

Certificate Number: AT-3317



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Millstone Medical Testing - CT, LLC**

35 Griffin Rd South  
Bloomfield, CT. 06002

Victoria Hughes 508-728-5608  
[victoria.hughes@millstonemedical.com](mailto:victoria.hughes@millstonemedical.com)

**TESTING**

ISO/IEC 17025 Accreditation Granted: **14 November 2025**

Certificate Number: **AT-3317** Certificate Expiry Date: **18 October 2026**

**Biological**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Sterility Test for Biological Indicators/Paper Strips	LSPEC-0001 (M100) USP<55> ANSI/AAMI/ISO 11138	Medical Devices	ISO 5 Laminar Flow Hood Incubators
Product Sterility	LSPEC-0003 (M150) USP <71>, ISO 11135, ISO 11138, ANSI/AAMI/ISO 11737-2, ANSI/AAMI/ISO 11137-2, ANSI/AAMI/ISO 13004	Medical Devices	ISO 7 Clean Room ISO 5 Biological Safety Cabinet Incubators
Total Bioburden	LSPEC-0004 (M200) LSPEC-0005 (M200P) ANSI/AAMI/ISO 11737-1	Medical Devices	ISO 5 Laminar Flow Hood Incubators
Bioburden Recovery (Inoculated Product Method)	LSPEC-0006 (M210) ANSI/AAMI/ISO 11737-1	Medical Devices	ISO 5 Laminar Flow Hood Incubators
Bacteriostasis and Fungistasis Test	LSPEC-0017 (M400) USP <71> ANSI/AAMI/ISO 11737-2	Medical Devices	ISO 5 Laminar Flow Hood Incubators
Cytotoxicity Extract/Elution Test (ISO/USP)	LSPEC-0043 (T150) ISO 10993-5, ISO 10993-12 USP <87>, USP <88>	Medical Devices	ISO 5 Laminar Flow Hood MEM Cells Microscope CO <sub>2</sub> Incubator



ANSI National Accreditation Board

**Biological**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Bacterial Endotoxin Gel clot method	LSPEC-0022 (M407) ANSI/AAMI ST72 USP <161> USP <85>	Medical Devices	Lysate CSE Heat Block
Bacterial Endotoxin (Kinetic Chromogenic)	LSPEC-0046 USP <85>, USP <161>, ANSI/AAMI ST72	Medical Devices	Nexgen-PTS Reader Laminar Flow Hood Depyrogenation Oven
Inhibition & Enhancement	LSPEC-0022 (M407) LSPEC-0046 USP <85>, USP <161>, ANSI/AAMI ST72	Medical Devices	Lysate CSE Heat Block Nexgen-PTS Reader ISO 5 Laminar Flow Hood Depyrogenation Oven

**Chemical**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Total Organic Carbon	LSPEC-0047 USP <643>	Medical Devices Water	ISO 5 Laminar Flow Hood TOC Analyzer Incubator Sonicator

Jason Stine, Vice President